



Sterilizing Metal Impression Trays

After removing any remaining impression material, the tray should be processed through an ultrasonic cleaner or thermal disinfectant, then dried, placed into a sterilization pouch, and processed through a Chemiclave or dental autoclave (*typically operating at 250° or 270°F*).